



Ref. Certif. No.

DE 3 - 32670

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Electronic measuring equipment
Name and address of the applicant	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19 74385 Pleidelsheim GERMANY
Name and address of the manufacturer	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Name and address of the factory	E. ZOLLER GmbH & Co. KG Gottlieb-Daimler-Str. 19, 74385 Pleidelsheim, GERMANY
Ratings and principal characteristics	see next page
Trademark / Brand (if any)	ZOLLER
Customer's Testing Facility (CTF) Stage used	CTF STAGE 1
Model/type Ref.	genius, threadCheck, edgeControl, titan, 3D Check
Additional information (if necessary)	Certificate DE 3 - 32064 issued 2023-05-15 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	713368057

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CB 095184 0032 Rev. 02

Date, 2025-09-25

Siemon

(Thorsten Siemon)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service



Ref. Certif. No.

DE 3 - 32670

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

Ratings and principal characteristics

Rated Voltage: 100-120 / 200-240 V
Rated Frequency: 50/60 Hz
Rated Power: 600 VA
Protection Class: I
Ambient Temp.: Restricted: 5 to 35 °C, 5-90 % r.H. (non-condensing)
Air Pressure: 6-8 bar
Pollution Degree: 2
Overvoltage Category: II

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CB 095184 0032 Rev. 02

Date, 2025-09-25

Siemon

(Thorsten Siemon)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service